



## NITRILE, POWDER-FREE, SINGLE USE EXAMINATION GLOVE

High performance, powder-free, nitrile single use examination glove. Significantly stronger alternative to latex and food safe.

### FEATURES

- › Comfortable with a 'second skin' feel
- › High resistance to penetration and projections of many chemicals
- › Textured finish on fingertips for improved grip and touch
- › Powder-free, non-sterile
- › AQL 1.5
- › Approved for food handling
- › Latex free - ideal for people with a type I latex sensitivity
- › Conforms to EN455 (all parts), ASTM D6319-10 and D6124-06
- › Medical device compliant (MDD 93/42/EEC)
- › Thickness: 0.10mm (5g)



### SUITABLE FOR

#### Typical Industries

- Aerospace
- Chemical
- Electronics
- Food
- Pharmaceutical
- Manufacturing
- Oil and Gas

#### Suitable Applications

- Chemical Handling
- Food Handling
- Handling Oily Components
- Laboratory Analysis
- Light Assembly
- Paint Spraying

### CERTIFICATION

CAT III	EN ISO 374-1:2016/ Type B	EN ISO 374-5:2016		
2777	KTP	Virus		MEDICAL DEVICE

See overleaf for explanation



## PRODUCTION INFORMATION

MATERIALS	LINER:	Nitrile
	COATING:	–
COLOUR	Blue	
LENGTH (mm)	240	
CUFF STYLE	Beaded	

## ORDERING DETAILS

SIZE	CODE	PACKAGING
6½-7/S	SKYUTAH1	100 singles per box
7½-8/M	SKYUTAH2	
8½-9/L	SKYUTAH3	
9½-10/XL	SKYUTAH4	
10½ - 11/XXL	SKYUTAH5	10 boxes per case

## RECOMMENDATIONS FOR USE

- **USE:** Single use glove. Not suitable for thermal, electrical, heavy chemical protection. Do not use near moving machines if there is a risk of entanglement
- **STORAGE:** Store in dry conditions in the original packaging and away from direct sunlight
- **CLEANING:** To clean, wipe with a damp cloth. Note: The performance characteristics of worn and laundered gloves may differ from the results shown. Inspect the gloves to ensure no damage is present
- **LIFETIME:** Service life depends on the glove application and therefore cannot be specified. It is the responsibility of user to ensure the glove is suitable for its intended use

## CERTIFICATION LEGENDS



PROTECTION AGAINST  
MICRO-ORGANISMS EN 374-5  
VIRUS = Glove has passed ISO 16604:2004 (method B)



FOOD CONTACT  
DIRECTIVE 2002 / 72 EC



RESISTANCE TO CHEMICAL PERMEATION - EN ISO 374:2016

CODE	CHEMICAL	CODE	CHEMICAL	TYPE OF GLOVES	BREAKTHROUGH TIME
A	Methanol	J	n-Heptane	A	≥30 min for at least 6 chemicals
B	Acetone	K	Sodium hydroxide 40%		
C	Acetonitrile	L	Sulphuric acid 96%		
D	Dichloromethane	M	65% Nitric acid	B	≥30 min for at least 3 chemicals
E	Carbon Disulfide	N	99% Acetic acid		
F	Toluene	O	25% Ammonium hydroxide	C	≥10 min for at least 1 chemical
G	Diethylamine	P	30% Hydrogen peroxide		
H	Tetrahydrofurane	S	40% Hydrofluoric acid		
I	Ethyl acetate	T	37% Formaldehyde		

## GLOBUS

### EUROPE

[www.globusgroup.com](http://www.globusgroup.com)

**E:** [sales@globus.co.uk](mailto:sales@globus.co.uk)

**T:** +44 (0)161 877 4747

**F:** +44 (0)161 877 4746

### MIDDLE EAST & AFRICA

[www.globusgroup.com/gcc](http://www.globusgroup.com/gcc)

**E:** [gcc@globusgroup.com](mailto:gcc@globusgroup.com)

**T:** +971 4 882 9962

**F:** +971 4 882 9963

### GLOBUS AMERICAS

[www.globusgroup.com](http://www.globusgroup.com)

**E:** [americas@globusgroup.com](mailto:americas@globusgroup.com)

Globus Group, T2 Trafford Point, Twining Road, Trafford Park, Manchester, M17 1SH, UNITED KINGDOM  
 Globus EMEA FZE, Jafza One, Tower A, Office 2201, Jebel Ali, PO Box 61195, Dubai, UNITED ARAB EMIRATES  
 Globus Americas LLC, 45 Rockefeller Plaza, Suite 2000, New York City, New York, 10111, USA

EU Type-Examination Certificate issued by SATRA Technology Europe Limited, Bracetown Business Park, Clonee, D15YN2P, Republic of Ireland. (Notified Body No. 2777)